公司简介 Company Profile

伏羲生物科技(广东)有限公司 CareAble Biotechnology Co., Ltd

Welcome to CAREABLE



中国医药保健品进出口商会非医用口罩白名单企业





中国医药保健品进出口商会

服务产业链!助力国际化

首页

关于商会 ~

Careable Biotechnology Co., Ltd.

新闻中心▼

行业服务 -

权威发布 ▼

闻中心 > 通知公告

动态更新:取得国外标准认证或注册的非医用口罩生产企业清单

2020年06月12日 中国医药保健品进出口商会

	取得国外标准认证或注册的非医用口罩生产企业清单								
	动态更新 日期: 2020年6月12日 下载								
序	生产企业	统一社会信用代码	国外注册						
号			认证情况						
151		91440700MA54BFN314	欧盟CE						

双CE认证: NB2163 B&C2 NB2797 B&C2

简介 About us

公司介绍:

伏羲生物料科技(广东)有限公司

CareAble Biotechnology Co., Ltd

是一家专业的民用防护口罩生产企业。我们拥有专业的技术人员,先进的生产设备和监控测量设备。我们的产品:

CARE002 FFP2防护口罩取得欧盟NB2163, Module B+C2 CE 认证 CARE003 FFP2防护口罩取得欧盟NB2797, Module B+C2 CE 认证 CARE006 KN95防护口罩取得GB2626-2019 新国标认证.

我们目前拥有多条口罩生产线,严格把控每批次的口罩的品质,坚持质量至上,为广大客户提供优质的产品和服务,欢迎全球客商前来洽谈合作。









产品图片/Pictures

我们的产品 Our Product


CARE002 FFP2



CARE003 FFP2



CARE006 KN95



营业执照 Business License



91440700MA54BFN314

营业执照

(副本)(副本号:1-1)

称 伏羲生物科技 (广东) 有限公司 型 有限责任公司(自然人投资或控股)

法定代表人 谢英

经 营 范 围 生物技术研发及推广;研发、生产、销售;医护产品,防护用品,劳保用品,医疗器械,环保产品,消毒用品(不含危险化学品),知识产权服务,版权代理服务,货物或技术进出口(国家禁止或或涉及行政事推的货物和技术进出口降外)。(依法须经批准的项目,经相关部门批准后方可开展经营活动。)■

注册资本 人民币伍佰万元

成立日期 2020年02月20日

营业期限长期

住 所 江门市宏兴路3号0幢(自編002) (信息申报制)

登记机关

2020 年 3 月 25

国家企业信用信息公示系统网址: http://www.gsxt.gov.cn

市场主体应当于每年 1月1日 至 6月30日通过 国家企业信用信息公示系统报送公示年度报告 国家市场监督管理总局监制

扫描二维码登录" 国家企业信用信息 公示系统"了解更 多登记、备案、许 可、监管信息。

CARE002

CE 2163 Module





Certificate No: 2163 - PPE - 677

Respiratory protective devices, filtering half masks to protect against particles manufactured by

Building O, 3rd Hongxin Road, Jiangmen City, Guangdong, China are tested and avalers.

EN 149:2001 + A1:2009 Respiratory Protective Devices -Filtering Half Masks to Protect Against Particles -Requirements, Testing, Marking

Based on the type examination conducted with the evaluation of test reports, technical file according to Personal Protective Equipment Regulation (EU) 2016/425 Annex 5, it is approved that the product meets the requirements of the regulation.

Product Definition

Model: CARE002 Filtering half mask Total Inwards Leakage: Class - FFP2

Here by the manufacturer is allowed to use notified body number (2163) and can fix CE mark, as shown below, on the Category III product models given above, with;

- Issuing an appropriate EU Declaration of Conformity according to Personal Protective Equipment Regulation (EU) 2016/425 Annex 9.
- Ongoing successful performance in fulfilment of the requirements set out in Personal Protective Equipment Regulation (EU) 2016/425 and harmonised standards, ensured by assessments based on Annex 7 (Module C2) or Annex 8 (Module D) of the regulation no later than I year from the beginning of serial production

This certificate is initially issued on 12 / 05 /2020 and will be valid for 5 years, if there is no change in the relevant harmonised standard affecting the essential health and safety requirements.

2163

Suat KACMAZ UNIVERSAL CERTIFICATION Director

CARE002

CE 2163 Module C2





CERTIFICATE OF CONFORMANCE

Certificate No: 2163 - PPE - 677/01

Respiratory protective devices, filtering half masks to protect against particles manufactured by

CAREABLE BIOTECHNOLOGY CO., LTD.

Building O, 3rd Hongxin Road, Jiangmen City, Guangdong, China Continues to fulfil the requirements of

EN 149:2001 + A1:2009 Respiratory Protective Devices Filtering Half Masks to Protect Against Particles Requirements, Testing, Marking

Based on the evaluation of test reports and internal quality control audit reports according to EN 149+A1:2009 and Personal Protective Equipment Regulation (EU) 2016/425 Annex VII (Module C2). This certificate implies that the manufactured products show below are in conformance with the approved EU Type Examination model and meets the requirements of the regulation.

Product Definition

Model查验	Class	EU Type Examination Certificate				
上连消失	Class	Serial No.	Date	Issuing NB Nr.		
CARE002	FFP2	2163-PPE-677	12.05,2020	2163		

Here by the manufacturer is allowed to use notified body number (2163) and can fix CE mark, as shown below, on the Category III product models given above, with;

- Issuing an appropriate EU Declaration of Conformity according to Personal Protective Equipment Regulation (EU) 2016/425 Annex 9.
- Taking all measures necessary so that the manufacturing process and its monitoring ensure the homogeneity of production and conformity of the manufactured PPE with the type described in the EU type examination certificate.

This certificate is issued on 12.05./2020 and will be valid for one year, until 11/05/2021 if the manufacturer makes no major change in the product designs and manufacturing processes affecting the product performance on the essential health and safety requirement.



Outten 3

Suat KAÇMAZ
UNIVERSAL CERTIFICATION
Director

> CARE002 EN149测试报告 EN149: 2001+A1:2009





检验检测报告

TEST REPORT



STFCE20200048

Product Name	CAREABLE Filtration Respirators	_
Trust Unit	Careable Biotechnology Co., Ltd	
Manufacturer		
Test Category	Entrusted Inspection	



Test Report

Product Name	CARPADI P. Pile di P.	Specification Type	CARE002
Product Name	CAREABLE Filtration Respirators	Trademark	以子——
Trust Unit	Careable Biotechnology Co., Ltd	Tel	11/1/1/1/1/1/1/1/1/1/1/1/1/1/1/1/1/1/1/1
Manufacturer	***	Sample Grade	FFP2
Sample Quantity	70 4 1	Sample Receiving Date	2020-04-14
Test Category	Entrusted inspection	Serial Number	18 - 18 P
Samples Conditions	Meet the testing requirements	ARI. XX	**
Document and Decide Accordance	EN 149: 2001+A1: 2009 Respirate particles-Requirements, testing, mar		ltering half masks to protect again
1	The samples were tested, the is standard for FFP2 level.	-4/5 X	alc: \2020-0430
Test Conclusion	W.T ITO X	Signature D	超短短路 414十

Approver



Examine



Major tester





STFCE20200048

page 2 of 10

7.5 Material Pass¹

Materials used shall be suitable to withstand handling and wear over the period for which the particle filtering half mask is designed to be used.

Any material from the filter media released by the air flow through the filter shall not constitute a hazard or nuisance for the wearer.

After undergoing the conditioning described in 8.3.1 none of the particle filtering half masks shall have suffered mechanical failure of the facepiece or straps,

When conditioned in accordance with 8.3.1 and 8.3.2 the particle filtering half mask shall not collapse.

Note2: Refer to Annex A for test data.

7.6 Cleaning and disinfecting

N/A2

If the particle filtering half mask is designed to be re-usable, the materials used shall withstand the cleaning and disinfecting agents and procedures to be specified by the manufacturer.

Note3: Non-reusable respirator.

7.7 Practical performance

D3

The particle filtering half mask shall undergo practical performance tests under realistic conditions.

Note4: Refer to Annex A for test data.

7.8 Finish of parts

Pass

Parts of the device likely to come into contact with the wearer shall have no sharp edges or burrs.

STFCE20200048

page 3 of 10

7.9.1 Total inward leakage

D4

For particle filtering half masks fitted in accordance with the manufacturer's information, at least 46 out of the 50 individual exercise results (i.e. 10 subjects x 5 exercises) for total inward leakage shall be not greater than:

and, in addition, at least 8 out of the 10 individual wearer arithmetic means for the total inward leakage shall be not greater than

22% for FFP1, 8% for FFP2, 2% for FFP3

Note5: Refer to Annex A for test data.

7.9.2 Penetration of filter material

Page 5

The penetration of the filter of the particle filtering half mask shall meet the requirements of Table 1.

	Sodium chloride test 95	Paraffin oil test 95 1/min
FFP1	≤20%	≤20%
FFP2	≤6%	≤6%
FFP3	_ ≤1%	≤1%

This report may not be published except in full unless permission for the publication of an approved extract has been obtained in writing.

Note6: Refer to Annex A for test data.

7.10 Compatibility with skin

Pass⁶

Materials that may come into contact with the wearer's skin shall not be known to be likely to cause irritation or any other adverse effect to health.

Note7: Refer to Annex A for test data.

7.11 Flammability

Page⁷

When tested, the particle filtering half mask shall not burn or not to continue to burn for more than 5 s after removal from the flame.

Note8: Refer to Annex A for test data.

7.12 Carbon dioxide content of the inhalation air

Pagg⁸

The carbon dioxide content of the inhalation air (dead space) shall not exceed an Note9: Refer to Annex A for test data.

STFCE20200048

page 4 of 1

7.13 Head harness

Page⁹

The head harness shall be designed so that the particle filtering half mask can be donned and removed easily.

The head harness shall be adjustable or self-adjusting and shall be sufficiently robust to hold the particle filtering half mask firmly in position and be capable of maintaining Note10: Refer to Annex A for test data.

7.14 Field of vision

10

The field of vision is acceptable if determined so in practical performance tests.

Notell: Refer to Annex A for test data.

7.15 Exhalation valve

M/A II

A particle filtering half mask may have one or more exhalation valve(s), which shall function correctly in all orientations.

If an exhalation valve is provided it shall be protected against or be resistant to dirt and mechanical damage and may be shrouded or may include any other device that may be necessary for the particle filtering half mask to comply with 7.9.

Exhalation valve(s), if fitted, shall continue to operate correctly after a continuous exhalation flow of 300 l/min over a period of 30 s.

When the exhalation valve housing is attached to the faceblank, it shall withstand axially a tensile force of 10 N applied for 10 s.

Note12: Valve-less respirator.

7.16 Breathing resistance

- ×12

720	Maximum permitted resistance (mbar)						
Classification	Inha	Exhalation					
12 2	30 l/min	95 l/min	160 l/min				
FFP1	0.6	2.1	3.0				
FFP2	0.7	2.4	3.0				
FFP3	1.0	3.0	3.0				

Note13: Refer to Annex A for test data.

STFCE20200048

page 5 of 10

7.17 Clogging

NI/A 13

7.17.2 Breathing resistance

14 13

Valved particle filtering half masks:

After clogging the inhalation resistances shall not exceed:

FFP1: 4 mbar, FFP2: 5 mbar, FFP3: 7 mbar at 95L/min continuous flow

The exhalation resistance shall not exceed 3 mbar at 160 L/min continuous flow

Valveless particle filtering half masks

After clogging the inhalation and exhalation resistances shall not exceed:

FFP1: 3 mbar, FFP2: 4 mbar, FFP3: 5 mbar at 95L/min continuous flow

7.17.3 Penetration of filter material

5° 1867	Sodium chloride test 95	Paraffin oil test 95 l/min
FFP1	≤20%	≤20%
FFP2	≤6%	≤6%
FFP3	≤1%	≤1%

N/A 13

Note14: Non-reusable respirator.

7.18 Demountable parts

NI/A 14

All demountable parts (if fitted) shall be readily connected and secured, where possible by hand

Note15: No demountable parts.

STFCE20200048

page 6 of 10

Clause	NO AZ		Result	Assessmen
2	Simulated	1# 35	No mechanical failure	160
11/6	wearing	20	No mechanical failure	
Mater	treatment	3#	No mechanical failure	n.
Mater	36 18 190	40	No mechanical failure	Pass
1	Temperature conditioned	5#	No mechanical failure	
1-7	conditioned	6#	No mechanical failure	
Practi	ical	7#	No mechanical failure	No. 3
perform	As received	80	No mechanical failure	Pass
X	AN THE	In	dividual exercise result	N. S.
219	13	9#	47 out of the 50 individual exercise results ≤ 11%	
1	XXX	10#	47 out of the 50 individual exercise results ≤11%	
8	As received	11#	47 out of the 50 individual exercise results ≤ 11%	
·V	7 230	12#	47 out of the 50 individual exercise results ≤11%	
110	V. TO)	13#	47 out of the 50 individual exercise results ≤ 11%	
	12 T	14#	47 out of the 50 individual exercise results ≤11%	
	WILL STATE	15#	47 out of the 50 individual exercise results ≤ 11%	
LXX	Temperature	16#	47 out of the 50 individual exercise results ≤11%	
24	conditioned	17#	47 out of the 50 individual exercise results ≤ 11%	
Total in	ward	18#	47 out of the 50 individual exercise results ≤11%	1. XXX.
leaka	ge de	Individ	ual wearer arithmetic means	Pass
1.10	2 825	9#	9 individual wearer arithmetic means≤ 8%	
1	* 2	10#	9 individual wearer arithmetic means≤ 8%	
1	As received	11#	9 individual wearer arithmetic means≤ 8%	15,00
31	XX VET	12#	9 individual wearer arithmetic means≤ 8%	
4	The state of the s	13#	9 individual wearer arithmetic means≤ 8%	
Mr.	1/2/	14#	9 individual wearer arithmetic means≤ 8%	
1	3/67	15#	9 individual wearer arithmetic means≤ 8%	
1 3	Temperature	16#	9 individual wearer arithmetic means≤ 8%	
1/2 S	conditioned	17#	9 individual wearer arithmetic means≤ 8%	
1-1	· X . /	18#	9 individual wearer arithmetic means< 8%	

CARE002 EN149测试报告 EN149: 2001+A1:2009

STFCE20200048

page 7 of 10

Clause		Z V82	Xc	Result	Assessmen	
	11/2	- V.79/	Sodi	ium chloride test(95L/min)	14.	
	A	然ケー	19#	0.81		
	356	As received	20#	0.72		
	26.35	31	21#	0.76		
	1-4	Simulated	22#	0.89		
	X	wearing	23#	0.94		
	A IZ	treatment	24#	0.97		
	K Illin	XX	25#	1.23		
	2791	M.S.+T.C.	26#	1.31		
	Penetration	X	27#	1.26	200	
7.9.2	of filter material/%	_	Pa	araffin oil test(95L/min)	Pass	
	material/%	1,07	28#	1.68		
	Mr.	As received	29#	1.74		
	· de	%	30#	1.79		
	- 20	Simulated	31#	1.88		
	~ XXX ,	wearing	32#	1.96		
	124.	treatment	33#	1.92		
	The same	28	34#	2.04		
	NEW	M.S.+T.C.	35# 💸	2.18		
	No. No.	XX	36#	2.11	1	
	:32	20,	9#	No irritation or any other adverse effect to health	121	
	1	-/	10#	No irritation or any other adverse effect to health		
	TINK.	As received	11#	No irritation or any other adverse effect to health	2 9	
	2	K Hill	12#	No irritation or any other adverse effect to health	1	
= 40	Compatibility	279	13#	No irritation or any other adverse effect to health	-1155	
7.10	with skin	7 1	14#	No irritation or any other adverse effect to health	Pass	
	34XX	18	15#	No irritation or any other adverse effect to health		
	XXX	Temperature	16#	No irritation or any other adverse effect to health		
	1700	conditioned	17#	No irritation or any other adverse effect to health		
	180	11.3	18#	No irritation or any other adverse effect to health	XXX.	
L	NE.	x - 30	37#	Didn't burn	7	
	XX VIV	As received	38#	Didn't burn	1	
7.11	Flammability	Temperature	39#	Didn't burn	Pass	
		conditioned	40#	Didn't burn	~ Digit	

CARE002 EN149测试报告 EN149: 2001+A1:2009

STFCE20200048

page 8 of 10

Clause		SK.		Res	sult	III)	Assessment
	Carbon		N	As re	ceived	- 100	The same
	dioxide	41#	20	42#	43#	Mean value	
7.12	content of the inhalation air/%	0.52	~ 1	0.53	0.51	0.52	Pass
54	15/25	.v	7	As rec	peived	NX NZ	-16
16	KT	9#	_16	Head harness can be donn sufficiently robust to h	1 11/11	1.19//	
	***	10#	*	Head harness can be donn sufficiently robust to h	2001	36. 1 38	
	N. IK	11#		Head harness can be donn sufficiently robust to h	/ /	1. 1. 1. O.	
	WY	12#	3	Head harness can be donn sufficiently robust to h	1	N 108 V	
C THE	Head	13#	W.	Head harness can be donn sufficiently robust to h	11.1		Pass
7.13	hardness	VY	X 400				
	NE.	14#	1	Head harness can be donn sufficiently robust to h	1100	1001	
	X TO THE REAL PROPERTY.	15#	8	Head harness can be donn sufficiently robust to h	Y/\\ . X(t)	111	
	AIX	16#	31.	Head harness can be donn sufficiently robust to h	//	1	TOWN /
	III.	17#	8	Head harness can be donn sufficiently robust to h	- 1	. 756 \	
E THELL		18#	W.	Head harness can be donn sufficiently robust to h	-7/10	1	Mr. 3
7.14	Field of vision	As	7#	Passed the	e practical performanc	e tests	Pass
LUI	2.000 01 7131011	received	8#	Passed the	e practical performanc	e tests	1 400

CARE002 EN149测试报告 EN149: 2001+A1:2009

STFCE20200048

page 9 of 10

Clause	1		20	Result	Y	Assessmo
W.7	-	10/	Inhala	ation	Exhalation	
11.0	1.3%	-	30 1/min	95 I/min	160 l/min	25
	J-201		X	As received	14	V.T
	25	A	0.3	0.9	1.4	III.
12	' '	В	0.3	0.9	1.4	
X.	41#	C	- 0.3	0.9	V 1.4 Z	100
	_ 1000	D	0.3	- 7 0.9	1.4	12
	(3)	E	0.4	1.0	1.5	1
	' X	A	0.3	0.9	X 1.4	()
	15	В	0.3	0.9	1.4	05
	42#	С	0.4	0.9	1.5	100
M	_	D	0.3	1.0	1.4	X
1K	X	E	0.3	0.9	1.4	S 1
A	3827	A	0.3	V /0.9	1.4	X.
-	%	В	0.4	0.9	1.5	WI
	43#	c	0.3	1.0	1.4	Dr.
Breathin	- I V	D	0.3	0.9	1.4	n -7
resistan		E	0.3	0.9	1.4	Pass
(iii bar)	5	180	Simulate	d wearing treatmen	it X	X
	7 3	A	0.3	0.9	1.4	-7
	XX	В	0.3	0.9	1.4	. 17.
- 20	44#	C	0.4	1.0	1.4	× 11/13
115	c 1	D	0.3	0.9	1.5	2701
"21	1	E	0.3	0.9	1.4	7
	10	A	0.4	1.0	1.5	W.
	2	В	0.3	0.9	1.4	X
	45#	C/	0.3	0.9	1.4	V.T.
× 350	35	D	0.3	- 0.9	1.4	The same
4	W	E	0.3	0.9	1.4	x-28
	Michigan	A	0.3	1 0.9 X	1.4	TXX.
		В	0.3	0.9	1.4	4
K Illist	46#	C	0.3	0.9	26 T 1.4	
10/	12	D	0.4	1.0	1.5	.17.
	N	E	0.3	0.9	1,4	~ 1/27



STFCE20200048

page 10 of 10

Clause		Result				Assessmen
Y 1		(3)	Inha	lation	Exhalation	(1)
AL THE	36	Z.	30 l/min	95 l/min	160 I/min	1 6
W. W.	N Vin		Temp	perature conditioned	K, ABD.	1. J. O.
XXX V	7	A	0.3	0.9	1.4	11/25
-7	1	В	0.3	0.9	4 1.4	-10
170	47#	C	0.4	1.0	X 1.5 X	5
700	X-TO	D	0.4	0.9	1.5	11/2
10/ X	3	E	0.3	1.0	1.4	-no
7.16 Breathing		A	0.3	0.9	2 1.4	(a)
resistance		В	0.3	0.9	1.4	Pass
7	48#	С	-0.4	1.0	1.4	12
X.T	1	D	0.3	0.9	1.5	XX
110	1	E	0.4	1.0	1.5	X
1 TO	IN	A	0.3	0.9	1.4 VT	150
18	1/4	В	0.4	1.0	1.5	1
4.	49#	C	0.3	0.9	1.4	26
The.	1 3	D	0.3	0.9	1.4	31
	-4/2	E	-0.3	0.9	1.4 /2/2	100 X
7.16 Breathing resistance	B: faci C: faci D: lyin	ng verti ng verti g on the	ctly ahead ically upwards ically downwards e left side e right side	(本)	此文件、香源	THE STATE OF THE S

Remarks : M.S.: Mechanical strength; T.C.: Temperature conditioning

Original Sample



===== End of Report ====



JIANGSU QUALITY SUPERVISION AND INSPECTION CENTER FOR SPECIAL SAFETY PROTECTION PRODUCTS

CE2797-CARE003 Module B







EU Type Examination Certificate

This is to certify that:

Careable Biotechnology Co., Ltd 伏羲Careable Building 0, no.3, Hongxing road Jiangmen City (002) a Careable Guangdong China

Holds Certificate Number:

CE 728685

In respect of: eable

Model CARE003 Particulate Respirator To technical specification Annex II (EHSR) of the PPE Regulation (EU) 2016/425 PPE for use by healthcare professionals as per Commission recommendation 2020/403

on the basis that BSI carried out the relevant Type Examination procedures under the requirements with the Regulation (EU) 2016/425 of the European Parliament and Council relating to Personal Protective Equipment Regulation (PPE) Annex V (Module B) and meets the relevant health and safety requirements specified in Annex II

For and on behalf of BSI, a Notified Body for the above Regulation (Notified Body Number 2797):

Previous Notified Body: BSI 0086 First Issued: 2020-06-09

Latest Issue: 2020-06-09

Drs. Dave Hagerlaars, Managing Director

Effective Date: 2020-06-09 Expiry Date: 2021-06-09

Page: 1 of 3



making excellence a

nniber 3336-586, ill John M. Keynesplein 8, 1066 EP Westfordam, The Netherlands

CARE003 CE2797-Module B

文件不支持清关查验 EU Type Examination Certificate 此文件不支持清关查验

No. CE 728685

Product Specification

Product Name:

Product Type:

Particulate Respirator. Particulate filtering half masks for use by Healthcare professionals.

CAREOUS.

Classification:

FFP2 NR un-valved.

Technical Specification:

Technical specification to satisfy Annex II of the PPE Regulation (EU) 2016/425.

Product Description:

The respirator is non-reusable, secured to the face of the user by a pair of elasticated ear straps, and has no exhalation valve. The respirator is FFP2

dass, vertical fold flat type.

The respirator listed on this certificate is for use by healthcare workers, first responders and other personnel involved in the efforts to contain the COVID-19

virus and avoid its further spread.

The product covered by this certificate is not approved for industrial applications and the certificate is only valid as long as EU Commission recommendation sheet

2020/403 remains applicable.

Product Assessments:

BSI's PPE for Healthcare Professionals 2020/403 - RPE Technical Specification.

First Issued: 2020-06-09 Latest Issue: 2020-06-09

Expiry Date: 2021-06-09

Page: 2 of 3

This certificate has been issued by and remains the property of BSI Group The Nett 9, 1066 EP Amsterdam, The N 此文件不支持清 should be returned my religiously upon request.

To check/its validity telephone +31 20 3460780. An electronic certificate and be authenticated order.

851. Copin The Netherlands 8.V., registered in the Netherlands under number 3326426V, at John M. Keyn A interior of BSI Group of Companies. nghian 1066 EP Amsterdam, The Netherlan & Careable

CE2797-CARE003 Module C2







Conformity to Type based on Internal Production Control plus supervised product checks at random intervals

This is to certify that:

Careable Biotechnology Co., Ltd 伏羲Careable Building 0, no.3, Hongxing road Jiangmen City (002) 伏羲Careable Guangdong China

Holds Certificate Number:

CE 728686

In respect of: eable

For the manufacture of particulate respirators to technical specification to satisfy Annex II of the PPE Regulation (EU) 2016/425.

on the basis that BSI carried out the supervised production checks at random intervals under the requirements with the Regulation (EU) 2016/425 of the European Parliament and Council relating to Personal Protective Equipment Regulation (PPE) Annex VII (Module C2)

For and on behalf of BSI, a Notified Body for the above Regulation (Notified Body Number 2797):

Previous Notified Body: BSI 0086

First Issued: 2020-06-09 Latest Issue: 2020-06-09 Drs. Dave Hagerlaars, Managing Director

Effective Date: 2020-06-09 Expiry Date: 2021-06-09

Page: 1 of 3

伏羲Careable making excellence a habit



Careable This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keyne should be returned immediately upon request.

To check its velidity telephone +31 20 3160780. An electronic certificate can be authenticated online.

CARE003 CE2797-Module C2

伏羲Careable

Conformity to Type based on Internal Production Control plus supervised product checks at random intervals

No. CE 728686

Product manufactured by:

Careable Biotechnology Co., Ltd. 此文件不支持清关查验 Building 0, No. 3 Hongxing road Jiangmen City e Guangdong China

Product details

The respiratory protective device covered by the scope of this Module C2 Certificate and the Technical Specification to which the product is manufactured are as follows:

此文件不支持清关查验

Product type: able Particulate filtering half masks for use by Healthcare professionals.

CAREOUS FFP2 NR Model and classifications:

Technical specification to satisfy Annex II of the PPE Regulation (EU) 2016/425. **Technical Specification:** BSI's PPE for Healthcare Professionals 2020/403 RPE Technical Specification.

Latest Issue: 2020-06-09 Effective Date: 2020-06-09 Expiry Date: 2021-06-09

此文件不支持清关查验

Page: 2 of 3

This certificate has been assued by and remains the property of BSI Group The Netherlands B.V. John N. Keynesplein 9, 1066 EB Amsterdam, The should be returned immediately upon request.
To check its validity telephone + 31 20 9360780. An electronic certificate can be authenticated online.

der number 3325 1284, at John M. Keynespieln S, 1066 EP Amsterdam, The Nett

CARE003 BSI Testing Report EN149:2001+A1:2009

大義Careable 伏羲Careable 状義Careable 状義Careable 状義Careable

此文件不支持清关查验 此文件不支持清关查验

代義Careable 伏羲Careable 伏羲Careable 伏羲Careable 世文件不支持清关查验 此文件不支持清关查验 也文件不支持清关查验 A 大義Careable 大義Carea

Test Report 3204582. Test Report 3204582.

Careable Biotechnology Co., Ltd.

Careable HXCareable H areable 伏羲Careable 伏羲Careable 伏羲Careable 此文件不支持清关查验此文件不支持清关查验

查验 此文件不支持清关查验 de 伏羲Careable 伏羲Careable 伏羲Careable 伏羲Careable 此文件不支持清关查验 此文件不支持清关查验 Para Page Tofgareab' 供養Careable 供養Careab



Introduction.

3204582 - Test Report

This report has been prepared by Paul Waller and relates to the activity detailed below:

Job/Registration		
Job number:	3204582	Careable Biotechnology Co., Ltd.
Job type:	Testing Samples Submitted	Building 0, No. 3 Hongxing road
Start Date:	14/05/2020	Jiangmen City
Test type:	Type Tareable	Guangdong
Sample ID:	10189621 / 10190018	China ble
Registration:	CE 728685	area 大培清美堂和 伏敬
Scheme:	Negative pressure RPE	"文件不支"
Protocol:	PP123 A SECAYER	此文件
Scheme Manager:	Kinga Demetriou	ecareable 大在验 供養

The report has been approved for issue by T Wicksey - Senior Test Engineer

Approved For Issue		or areable		13
2014	此文件不又weable	小文件不支持	Careable	
持清美查验		Issue Date: 22 May 2020		TRX

Objectives.

This is an independent test evaluation to only certain clauses or sub-clauses of the agreed specification in accordance with the following test programme:
BSI COVID-19 filtering face piece technical specification, for COVID-19 masks for use by healthcare workers

Product Scope.

COVID-19 masks for use by healthcare workers

The samples were received on 11 May 2020 and the testing was started on 14 May 2020.

The samples submitted complied with the requirement.



文件不支持清关查验 3204582 - Test Report.

此文件不支持

areable

伏羲Careable 沃羲Careable

c支持清关查验 此文件不支持清关查验 es Test Samples.

est sa	mples.	e #XFT	丰不文心
Sample ID	ER Number	Description	Les Care
1 to 19	10189621 / 10190018	Model: CARE 003 FFP2 UN	

Description of Test Samples.

COVID-19 masks for use by healthcare workers: 伏羲Careable 伏羲Careable

Model: CARE 003 FFP2 UN 此文件不支持清关查验 支持清关查验



文件不支持清关查验 3204582 - Test Report. 此文件不支持清关查明

Test Requirements.

此文件不支持清关查验 大統 Careable Testing in accordance with BSI COVID-19 filtering face piece technical specification

Performance requirement	Test method clause	Requirement	Assessment
7.7 Practical performance The particle filtering half mask shall undergo practical performance tests	Testing shall be done in accordance with 8.4.	During the tests the particle filtering half mask shall be subjectively assessed by the wearer and after the	(文件) 伏羲Ci
under realistic conditions. These general tests serve the	eable 1X = ALX	test, comments on the following shall be recorded:	
ourpose of checking the equipment for imperfections that cannot be	此义	a) head harness comfort; b) security of fastenings;	
determined by the tests described elsewhere in this standard.	持清大學 伏羲Ca	c) field of vision; d) any other comments reported by	Pass
Where practical performance tests show the apparatus has	areable H	the wearer on request.	
mperfections related to wearer's acceptance, the test house shall	上培清关查验	Careable	
provide full details of those parts of the practical performance tests	Careable	此文件不支持用人 伏羲Care	
which revealed these imperfections.	Care	acareable 大歌	
? test subjects, masks tested 'As received'	不支持用人 Continue shall be diese in	All samples must achieve	able
7.9 Leakage 7.9.1 Total inward leakage	Testing shall be done in accordance with 8.5.	All individual exercise results tests shall be not greater than 11 % (for	
5 test subjects, masks tested 'As received'	*不支持清关鱼"	FFP2) and, in addition, all arithmetic means	Pass
大家 此文	acareable	for the total inward leakage shall be not greater than 8 % (for FFP2)	
7.9 Leakage 7.9.2 Penetration of filter	Testing shall be done in accordance with 8,11	6% for both PO and NaCl	查验
material 3 test samples masks tested 'As	- F 37 350 "	伏戰	Ca Pass
received', for NaCl (Sodium Chloride) and PO (Paraffin oil), Brilin test	大統Careable 伏羲Careable	验 此义"	
7.12 Carbon dioxide content of	Testing shall be done in accordance with 8.7.	The carbon dioxide content of the inhalation air (dead space) shall not	Careal
test samples, masks tested 'As received'	大概 Careabi	exceed an average of 1,0 % (by volume).	Pass
7.16 Breathing resistance	Testing shall be done in	The breathing resistances shall meet	清关查亚
3 test samples, masks tested 'As received'	accordance with 8.9	the requirements of; 30l/min – 0.7mbar (inhale)	伏羲Care
	ble Ka支持清	95l/min – 2.4mbar (inhale) 160l/min – 3.0mbar (exhale)	Pass
大歌 大歌	此文件	able 北文件不支	PER CO
Appendix A - Test Panel Data	伏羲	-ble	1/con
Product Photographs	eable - ##	持清大 山	地:惠关3



3204582 - Test Report.

Glossary of Terms.

Pass: Complies. Tested by BSI engineers at BSI laboratories

Pass 1: Complies. Witnessed by BSI engineers in manufacturers laboratory. Pass 2: Complies. Tests carried out by third party lab; results accepted by BSI.

Pass*: Report resulted in uncertainty and states that Compliance is more probable than non-compliance.

Fail: Non-compliance. Product does not meet the requirements of this clause.

Fail*: Report resulted in uncertainty and states that Non-compliance is more probable than compliance.

N/T: Not Tested N/A: Not Applicable AR: As Received

TC: Temperature Conditioned

SW: Simulated Wear FT: Flow Tested MS: Mechanical strength

MMDF: Manufactures Minimum Design Flow MMDC: Manufactures Minimum Design Condition

Conditions of Issue.

This Test Report is issued subject to the conditions stated in current issue of 'BSI Terms of Service'. The results contained herein apply only to the particular sample(s) tested and to the specific tests carried out, as detailed in this Test Report. The issuing of this Test Report does not indicate any measure of Approval, Certification, Supervision, Control or Surveillance by BSI of any product. No extract, abridgement or abstraction from a Test Report may be published or used to advertise a product without the written consent of BSI, who reserve the absolute right to agree or reject all or any of the details of any items or publicity for which consent may be sought.

Should you wish to speak with BSI in relation to this report, please contact Customer Services on +44 (0)8450 80 9000.

BSI

Kitemark House Maylands Avenue Hemel Hempstead Hertfordshire HP2 4SQ



Opinions and Interpretations expressed herein are outside the scope of our UKAS accreditation.

Unless otherwise stated, any results not obtained from testing in a BSI laboratory are outside the scope of our UKAS accreditation.

BSI Testing Report EN149:2001+A1:2009



3204582 - Test Report.

Test Results

Testing in accordance with BSI COVID-19 filtering face piece to

CLAUSE	REQUIREMENTS			ASSESSMENT
7.7	Practical performance	TIF X III	伏羲Car	此文件不
	The particle filtering half mask sh conditions. These general tests se imperfections that cannot be dete	erve the purpose of checking	the equipment for	
	Where practical performance test wearer's acceptance, the test hou performance tests which revealed	ise shall provide full details of		
	Test in accordance with clause 8.	4 of the standard.		Pass
	Testing in accordance with 8s specification, for masks for us During the tests the particle filter and after the test, comments on a) head harness comfort; b) secu comments reported by the weare	se by healthcare workers ing half mask shall be subjecthe following shall be record writy of fastenings; c) field of	s ctively assessed by the wearer led:	

Table A: Practical performance

Toch	-21	Comments area				
Test candidate	Sample	Head harness comfort	Security of fastenings	Field of vision	Any other comments	Assessment
JA1	1 AR	OK	OK C	Good	None (1880	Pass
JS3	2 AR	OK TA	OK	OK	None	Pass

7.9

7.9.1 Total inward leakage

The laboratory tests shall indicate that the particle filtering half mask can be used by the wearer to protect with high probability against the potential hazard to be expected.

The total inward leakage consists of three components: face seal leakage, exhalation valve leakage (if exhalation valve fitted) and filter penetration.

Test in accordance with clause 8.5 of the standard.

Testing in accordance with BSI COVID-19 filtering face piece technical specification, for masks for use by healthcare workers

5 test subjects, masks tested 'As received'. All individual exercise results tests shall be not greater than 11 % (for FFP2) and, in addition, all arithmetic means for the total inward leakage shall be not greater than 8 % (for FFP2).

14 K	互持用		veabl	6	Inward Leakag	e (%)	arear	5	半查验
Test	Sample	Pre test	Α	IL TBEA	C	D	E	医支持 /	
candidate	Sample	condition	Walking	Walking with head side to side	Walking with head up & down	Walking and talking	Walking	Average	Assessment
GR1	3.10	AR	0.13	0.11	0.17	0.05	0.23	0.14	Pass
LM2	4	AR	0.31	0.21	0.26	1.24	0.24	0.45	Pass
JS2	5	AR	2.74	2.98	2.48	0.70	1.89	2.16	Pass
JA1	6	AR	3.63	5.01	5.12	1.23	2.81	3.56	Pass
JB1	77	AR	1.87	2.54	2.11	3.21	3.24	2.60	Pass



CLAUSE REC	UIREMENTS		ASSESSMEN
7.9.2 Pen	etration of filter material	&Care 上共清关宣引	17.5%
	ting in accordance with BSI COV. cification, for masks for use by h	ID-19 filtering face piece technical	

Sample	Pre-test	11文件	Penetra	ation (%)	
number	condition	Flow through filter (I/min)	Limit	Actual	7
8	_ AR	有大		0.124	1
9	AR	hie 95	< 6	0.096	ab
10	AR a Y	San Try	14.	0.087	

Sample Pre-test		Flow through filter (I/min)	Penetra	ition (%)	
number	condition	rion dilough litter (() lillin)	Limit	Actual Ca	areas
11	AR	14大学	- 6/9	0.391	
12 6	AR	95	Care 6	0.602	
13	AR	大多		0.636	

7.12 Carbon dioxide content of inhalation air

The carbon dioxide content of the inhalation air (dead space) shall not exceed an average of 1.0% (by volume).

Campa b	Pre-test condition	Dead spa	ace CO ₂ (%)
Sample	Pre-test condition	Limit 17	Measured
.14	AR	reapic	0.43
· 清美型 15	AR	< 1.0	0.48
16, 16	AR	5特消入—	0.41
伏羲	此文件不	cable	业文件不



3204582 - Test Report.

Test Results. (Continued)

Coc Inc	continued)		
CLAUSE	REQUIREMENTS	able	ASSESSMENT
7.16	Breathing resistance	上海 关 但 。	17.82

Testing in accordance with BSI COVID-19 filtering face piece technical specification, for masks for use by healthcare workers

3 test samples masks tested 'As received'. Test in accordance with clause 8.9 of the

The breathing resistances shall meet the requirements of FFP2; 30l/min - 0.7mbar (inhale), 95l/min - 2.4mbar (inhale), 160l/min - 3.0mbar (exhale)

Sample	Pre-test	Continuous flow	Inhalation resis	tance (mbar)	
Sample	condition	(I/min)	Limit	Measured	
17	AR	· 本用	ble.	0.45	
b\918	AR	30	< 0.7	0.47	
19	AR		14. 不寸	0.45	
17	AR	eapie	ILIXIT	1.45	
18	AR	95	< 2.4	1.60	
ea 19	AR		- SECarea	1.50	

Table G: Clause 8.9 – Breathing resistance. Exhalation resistance at a continuous flow,

Sample condition (I/min) Limit Measured
The state of the s
120
18 AR 160 < 3.0 2.55
(2.37 上海 X) (2.37 上海 X) (2.37 上海 X)

Page 8 of gareable



文件不支持清关查验 3204582 - Test Report.

爱Careable 伏羲Careable

Test	代義(a)	TA	Facial Dimension	ons (mm)	THE	X Com - core
Candidate	Length of face	Width of face	Face depth	Width of mouth	Head Circumference	Sex Care
JA1	117	134	129	49	565 2610	Male
JS3	126	are 134	124	49	C 600	Male
GR1	124	145	126	49 (0	590	Male
LM2	110	148	125	44	589	Male
JS2	126	142 610	125	57 7 2	575 -020	Male
JB1	114	144	108	59	574	Male

Note: All candidates were clean shaven

Product photographs.



Front view





End of Report

[件不支持清关查验 ...making excellence a habit. Page 9 of 9 areable 支持清关查验

CARE006 国标报告 GB2626-2006

天纺标检测认证股份有阻公司 国家针织产品质量监督检验中心







检验报告









第1页 共4页

						Page I of 4
			(广东) 有限公司 technology Co., Lt	d	送样人: Contact	
l	委托单位/地址		technology Co., Lt 号0幢(自编002)	u	Contact	
l	Applicant	Building 0,3	rd, hongxing road,	jiangmen cit	y 电话: Tel.	
		guangdong Ch: 伏義生物科技	ina (广东) 有限公司			
ent	生产单位/地址	Careable Bio	technology Co., Lt	d		
Cli	Manufacturer		号0幢(自编002) rd, hongxing road,	iionamon ait	y guangdong Chin	
客户提供信息 Information Drovided by Client		样品名称:	一次性KN95防护口罩			IBIO
是 提 vide		Sample Name	KN95 FILTRATION I		Trademark	1010
提供信		样品总数:	90个			
信息	DI BI O di	Sample Count	90Pieces			
rmat	样品信息 Information of	号型规格:	CARE006		颜色: Colour	
Info	Submitted				安全类别:	
	Sample	质量等级: Quality Grade			Safety	
1		产品款号或货号:	50000000000000000000000000000000000000		Category	
1		「一面級写以页写: Style No. or				
		Order No.				
	样品描述	1# 4& DE III	W 1			
	Test Part Description	1# 白色口罩 Whit	e Mask			
	检验性质	委托检验	样品受理日期	0000 00 00	报告签发日期	9 0000 00 11
	Test Type	Commission Test	Date of Submission	2020-06-03	Date of Check	2020-06-11
					•	
То	执行标准 st Standards	GB 2626-2019 GB 2890-2009				
16	st Standards	GD 2090 2009			Ti di tim	船山
					KeT	THE PARTY OF THE P
1	检验结论	检验结果及符合性			渡一	麗
	Conclusion	Test results an	d compliance refer	to next page	(s). 写一	-22
1				检验单位盖音Sta	mp of Invocation U	ain &
非核	斥准检验方法说明			TEAST EMILTOUS	(e) 加湿
	andard Check Methods 合结果的不确定度					贴 样 Sample Attached
1,000	tainty of Result					зашрте літаслев
	备注		按GB 2626-2019标准		1 1	D
	Remarks	As per client s to the standard	request, the test GB 2626-2019.	t items are ju	aged according	Face side up
批准:		3. 四. 本 · 申	核: / /2	Mr.	编制:	ことろせ
Appro	ver	子子高 Ch	ecker W7	SIM	Editor	一时历

CARE006 国标报告 GB2626-2006

天纺标检测认证股份有阻公司 国家针织产品质量监督检验中心















第2页 共4页

TC-WT200356	29					Page 2 of
检测项目 Test Items	项目描述 Description	单位 Unit	标准值 Standard Requirement	实测值 Results	评价 Conclu sions	方法标准/备注 Test Method/ Remarks
1# 白色口罩	White Mask					
呼吸阻力 Respiratory	呼气阻力 Expiratory Resistance	Pa	≤210	未预处理样品 Samples Without Pretreatment 1: 112 2: 115 预处理样品 Samples With Pretreatment 1: 106 2: 103	符合	GB 2626-2019
Resistance	吸气阻力 Inspiratory Resistance	Pa	≤210	未预处理样品 Samples Without Pretreatment 1: 121 2: 127 预处理样品 Samples With Pretreatment 1: 113 2: 120	Pass	OB 2020 2019
可燃性 Flammability	续燃时间 Afterflame Time	S	≤5	0	符合 Pass	GB 2626-2019
	样品表面 Surface of The Sample		按标准要求 As per standard requirement	+		
外观检查 Appearance Inspection	部件材料和结构 Component Material and Construction		按标准要求 As per standard requirement	+	符合 Pass	GB 2626-2019
	部件经过温度湿度 预处理后 The components after the temperature and humidity pretreatment		按标准要求 As per standard requirement	+		
死腔 Dead Space		%	≤1	0.6	符合 Pass	GB 2626-2019
视野	下方视野 View Blow	0	≥35	61	符合	
View	双目视野 Binocular View	%	≥65	≥65	Pass	GB 2890-2009

CARE006 国标报告 GB2626-2006

天纺标检测认证股份有阻公司 国家针织产品质量监督检验中心















第3页 共4页 Page 3 of 4

CKTC-WT20035629)					Page 3 of 4
检测项目 Test Items	项目描述 Description	单位 Unit	标准值 Standard Requirement	实测值 Results	评价 Conclu sions	方法标准/备注 Test Method/ Remarks
头带 Head Harness			不应出现滑 脱、断裂。 Have no slippage or breaking.	10N负荷持续10秒,头带未出 现滑脱、断裂 The head harness have no slippage or breaking for 10s under 10N loading	符合 Pass	GB 2626-2019

表中"+"表示符合标准要求,"X"表示不符合标准要求。

以下空白 Blank Part Below

⁺ Meet the standard requirements, X Not Meet the standard requirements.

CARE006 国标报告 GB2626-2006

天纺标检测认证股份有阻公司 国家针织产品质量监督检验中心

检验报告













第4页 共4页

Page 4 of 4

CKTC-WT20035629

样 品 Sample



对外贸易经营者备案登记表 Foreign Trade Registration

对外贸易经营者备案登记表

备室登记表编号: 03818875

统一社会信用代码: 91·

91440700MA54BFN314

		15 17 1/2 1/2 1/2 1/2 1/2 1/2 1/2 1/2 1/2 1/2	
经营者中文名称	伏羲生物科技(广东)有限公司	
经营者英文名称	Careable Biotechi	nology Co., Ltd	
组织机构代码		经营者类型 (由备案登记机关填写	有限责任公司
住 所	江门市宏兴路3号() 幢 (自編002)	
经营场所 (中文)	江门市宏兴路3号() 幢 (自編002)	
经营场所 (英文)	Building O 3rd Hongs	kin Road Jiangmen City Gua	ingdong China
联系电话	0750-3089979	联系传真	
邮政编码	529000	电子邮箱	1196399272@qq. com
工商登记注册日期	2020-2-20	工商登记注册号	
The state of the s	AND THE PARTY OF T	Control of the Contro	

依法办理工商登记的企业还须填写以下内容

企业法定代表人姓名	谢英有效证件号	513024196103153984
注册资金	伍佰万元	(折美元)

依法办理工商登记的外国(地区)企业或个体工商户(独资经营者)还须填写以下内容

企业法定代表人/ 个体工商负责人姓名	有效	女证件号
企业资产/个人财产		(折美元)

备注		A	1			7	A	7	7	A	A	A	A	1	P.	A	7	
PS.																		×

填表前请认真阅读背面的条款、并由企业法定代表人或个体工商负责人签字、盖章。





企业海关备案编码 Customs Register code

海关进出口货物收发货人备案回执

企业名称	伏羲生物科技 (广东) 有限公司
统一社会信用代码	91440700MA54BFN314
海关备案日期	2020-03-26
海关编码	44079619Q6
检验检疫备案号	5952100040
有效期	长期




自然人、法人或者非法人组织可通过"中国海关企业进出口信用信息公示平台"(ht tp://credit.customs.gov.cn)或者"互联网+海关"(http://online.customs.gov.cn) 查询海关公示的企业信息。

品质控制 Quality Control

Filter efficiency Tester





		主控制	报表画
粒径	过滤效率%	当前压差 Pa	当前流量 L/min
≥0.3um	97. 93	141. 33	85. 09
≥0.5um	98. 521	设置时间s	測量计时 s
≥0.7um	98. 703	60	0,122,121,2
≥1.0um	98, 849	气溶胶浓度 mg/m³	
≥2.0um	99. 293	11. 402	
≥5. 0um	100.0	11.402	
气缸开	循环风开	测量启动	测量停止

品质控制 Quality Control

Inspection Instructions

Inspect instruction

ITEM No.	CARE002	Product	CAREABLE Filtration Respirators		File No.	FXQC-20005	FXQC-20005		
Admit		R&D Dept.		QA Dept.		TAB		Date	

INSPECT ITEM & STANDARAD

ITEM		Cassification & introduction	Inspect	Inspect standard		
ITEIVI		Specification & introduction	method/facility	AQL	LEVEL	
		Cleaning: Inspect at a distance of 50 cm within 3 seconds , not found any dirty spot, foreign particles, dust and hair.	Visual	0.65	N- II	
		2.Flat, no worn, no fold, no unfilled corner, and feel soft	Visual	0.4	N-II	
Mask	Appearance	3.Noseclip is totally integrated into the mask blank, and no any exposed, no sharp corner	Visual	0.4	N- II	
body		4. Ear loop is welded according the art work.		0 : OK 1:reject	5pieces/lot	
	Printing	The printing is clear and correct, no ghosting ,no missing.	Visual	0.65	N-II	
	Size	Overall size of mask : 155*105mm±5mm Nose bridge: 9±5mm	Ruler	0 : OK 1:reject	5pieces/lot	
	Appearance	The place of welding should be symmetry, and the inaccuracy should be little than 1mm. No worn, nocolor differenceand no dust.	visual/ruler	0.4	N-II	
Ear Loop	Length	Length: Left and right loop (weld spot exclude) 19cm ± 5mm		0 : OK 1:reject	5pieces/lot	
	Welding tension	Single spot's tension:10N and keep 10s. Double spots' tension:20N and keep 10s.	Fixed value weight	0 : OK 1:reject	10pieces/lot	
	Fabrics grams per square metre	Outter non-woven: 50gsm ±5% Melt-blown fabric: 25gsm ±5% Hot air filter cotton: 45gsm ±5% Inner non-woven: 25gsm ±5%	Electronic scale	0 : OK 1:reject	8pieces/lot	
Product Function	Hydrophilism	Water repellent(Outside ply)		0 : OK 1:reject	5pieces/lot	
	Respiratory	Water flow:95 L/min 1. inhale: <400Pa exhale: <400Pa	DMS-GL37	0 : OK 1:reject	9pieces/lot	
	resistance	2. inhale≤500Pa	DMS-GL37	0 : OK	12pieces/lot	

Inspect instruction

				1:reject		
Pe	enetrate	-01	DMS-GL36	0 : OK	9pieces/lot	
eff	ficiency	s6%	DMS-GL36	1:reject	spiecesnot	
Fil	iter	×94%	DMS-GL36	0 : OK	Spieces/lot	
eff	ficiency	25429	DMS-GL36	1:reject	opiedesnos	
Fit	lame	Resistance time ≤5s	Flame spread	0 : OK	delenented	
ret	tardancy	Resistance time ⇒ os	tester	1:reject	4pieces/lot	

Packing	1piecerbeg_20bags:box,50boxes/carton Packing is complete, products/quantity is correct. Boxes and cartons are no damaged&dirty, and packing according to the order.	Visual	2.5	N-II
Packing	Specification : blank disinfection bag Inner box :18.5*11*12.5cm±2mm carton : 5ply 57*39*64.5cm+1cm	Ruler	D : OK 1:reject	Spieces/lot
material	Printing&size should according to the artwork ,the lot number should meet the request of order.	Visual	0.4	N-II
	Box is qualified(no dirty &worn&deformation&unstuck&mosquito) No empty bag or box ,and all have stamped.	Visual	0.65	N-II

Attation: The item not specified is the same with sample.

Date	Change information	Cuatomer complaint	Mender

Explain: hand's fell and visual should be implement by veteran to guarantee product quality.

包装图 Packaging

CARE002 #1 Packaging:

package	pcs/ctn	CTN dimension (cm)			G.W/ctn	CBM per	
раскадс	pes/em	L W	Н	(kg)	ctn		
1pc per bag; 20 bags per box (18x10xH12.5cm); 50 boxes per carton	1000	57	38	64.5	13.00	0.140	



包装图 Packaging

CARE002 #2 Packaging:

package	pcs/ctn	CTN din		nsion	G.W/ctn	CBM per	
package	pcs/cm	L W H	(kg)	ctn			
5pcs per bag; 200bags per carton	1000	70	40	40	9.5	0.112	





包装图 Packaging

CARE003 #1 Packaging:

package	pcs/ctn CTN dimension (cm)			G.W/ctn	CBM per	
раскаде	L W	W	Н	(kg)	ctn	
1pc per bag; 20 bags per box (18x10xH12.5cm); 50 boxes per carton	1000	57	38	64.5	13.00	0.140





包装图 Packaging

CARE003 #2 Packaging:

package	pcs/ctn	CTN	dimer (cm)	nsion	G.W/ctn	CBM per	
раскаде	pcs/cui	L W F	Н	(kg)	ctn		
5pcs per bag; 200bags per carton	1000	70	40	40	9.5	0.112	





谢谢观看

THANK YOU

