

Certificate of Conformity

No.: EN190830006S

The following product has been tested by us with the listed standards and found in conformity with the European directive LVD 2014/35/EU. It is possible to use CE marking to demonstrate the conformity with this LVD Directive.

Applicant : NINGBO FENGHUA RUNDA MEDICAL APPARATUS

CO., LTD

Address : NO.137-1, SONGLIN WEST ROAD, SANGQIAO

INDUSTRIAL AREA, FENGHUA, NINGBO, CHINA

Manufacturer : NINGBO FENGHUA RUNDA MEDICAL APPARATUS

CO., LTD

Address : NO.137-1, SONGLIN WEST ROAD, SANGQIAO

INDUSTRIAL AREA, FENGHUA, NINGBO, CHINA

EUT : Drive System

Trade mark : N/A

M/N : RD-A398+ RD-398A+C01, RD-B398+ RD-398B+C02,

RD-D398+ RD-398C+C03, RD-E398+ RD-398D+C04, RD-F398+ RD-398G+C05, RD-E398+ RD-398Q+C06,

RD-E398+ RD-398E+C07

Technical Data : 100-240V~, 50/60Hz, 1.5A.

Test Standards : EN 60601-1: 2006+A1:2013+A12:2014



(Manager)
March 14, 2020

The certificate is based on a single evaluation of one sample of above-mentioned products. It does not imply an assessment of the whole production and does not permit the use of the test lab. Logo.

Ver.2.0